

envirosuite FS Form-003 2250 Based NMT PM Report

1. Site Information

1.1 CMS Ticket	393507			1.2 Airport/Urban	Airport
1.3 Inspection	Annual			1.4 Location	St. Doolaghs
1.5 Site Number	NMT02			1.6 Date	07-Sep-21
1.7 Inspected by	Kial Wallace/	David Courtney			
1.8 Normal NMT Operation Int	errupted	From	11:30:00	То	12:00

7. Occupational Health and Safety Check

Environmental exposure – Appropriate clothir	Yes		
Designated work are	nats, barricades as cordons	No	
Personal Protective Equipment (PPE) – protective glo	oves, safety boots, hi	gh viz vest, insect repellent	Yes
Animal hazards	No	Traffic hazards	No
Height / door work	No	Work within 2M of edge	No
Ladder used	Yes	Harness used	No
Pole erected	No	Support used	No
Multiple personnel required	Yes	Anchor points	No
Manual handling issues	Yes	Emergency equipment	No
Contact with electricity – Overhead power	No	Residual current devices	No
Structural earthing of buildings / installations	No	Risk of public injury	No
Working alone / isolated – means of communication	No	New hazards identified	No
Security or access issues	No		

Comments: Site Secure and no safety concern

11. Mechanical Inspection

i i. Mechanicai mapection						
Site Condition:		Good		✓		
Fence Condition:		Good		✓		
Cabinet Condition:		Good		√		
Moisture Inside NMT:		None		✓		
Insect Infestation:		None		✓		
Windshield Foam:		Replaced		√		
Bird Spike:		Good		✓		
Lightning Rod:		Good		✓		
Earth Connection:		Good		✓		
Pole and Pole Mounting Condition:		Good		✓		
Conduit Condition:		Good		✓		
Solar Panel Condition		NA				
Solar Battery Cabinet Condition:		NA				
Cleaned Solar Panel:		NA				
Battery Terminal Check:		Yes		√		
GPS Coordinates:	Altitude:	34 meters	Latitude:	53,41434	Longtitude:	-6,185997

40 Flootrical Imagestic

12. Electrical Inspection					
11.1 NMT LCD Display:	Good				
11.2 Door Switch:	Good				
1.11 NMT Model:	3639-A	1.9 NMT SNo:	3037367		
1.10 2250 SNo:	3008054	2250 Type:	G4 Light	11.5 Firmware Version:	4.5.0.92
	-	Cal Report No:		Lab Certificate Date:	
i		_			
Current Microphone Model:	4952	1.12 SNo:	3037367		
4 40 M T	0)//00				
1.13 Modem Type:	GX400	Modem SNo:		IMEI Number:	

Pistonphone/ Calibrator Information

2.1 Pistonphone/Calibrator Model:		B&K Type 4231 Calibrator			
2.1 Serial Number:	2460008	Cal Report No:	CDK2103317	2.2 Calibrated Date:	17/05/2021
2.5 Normalized Output (dB):	94,07	2.3 Atmos. Pressure (hPa)			
			2.4 Atmos	. Pressure Correction (dB)	

Backup Battery Check:

Battery voltage with Load & Mains (13 to 14V at -10°C to 40°C):

Battery voltage with load only, wait for 5 minutes before measuring (≥12V):

4			
1	2	3	4
14.4			
13.9			

Existing Microphone Calibration Acoustical Calibration:

Before Calibration

After Calibration

	_
2.6 Expected Calibration Level (dB): 94	Calibrating Level (dB): 93,86
Sensitivity (mV/Pa): 34,46	New Sensitivity (mV/Pa): 34,67
Deviation from initial (dB): 0,23	Deviation from last (dB): -0,1
Deviation from last (dB): 0,91	Deviation from initial (dB): 0,22
	CIC Calibration
	CIC Reference Ratio (dB) -36,71
	CIC Ratio (dB) -36,71
	CIC Result Passed
2250 Replacement (New) No	
2250 SNo:	
2250 Type: Cal Report N	lo : Lab Certificate Date:
<u> </u>	Firmware Version:
13. Microphone Replacement (New) No	
Microphone Model: SN	lo :
Acoustical Calibration:	
Before Calibration	After Calibration
Expected Calibration Level (dB):	Calibrating Level (dB):
Sensitivity (mV/Pa):	New Sensitivity (mV/Pa):
Deviation from initial (dB):	Deviation from last (dB):
Deviation from last (dB):	Deviation from initial (dB):
	<u> </u>
	CIC Calibration:
	CIC Reference Ratio (dB)
	CIC Ratio (dB)
	CIC Result
18. Weather Station N/A S	No: Confirm weather data:
<u> </u>	
Barix Instreamer	Confirm live data:
19. Final Check Completed Yes	
	<u> </u>
Any Other Comments:	
Windshield replaced	
Call EMS B&K Operation Centre (EMS OC) on 1-866-583 0	280 or +61 3 9508 4930 to verify NMT operation

2 of 2